

# TWL Calibration Data Summary

**Serial Number: 0301361**

**Full-Scale Flow Rate: 80**  
(yields 20mA output)

**Zero Flow Rate:** (yields 4mA output)

**Bridge Resistance:** 350.7 ohms

**Full-Scale Bridge Output:** 2.0423 mV/V

**Calibration Resistor (R7):** 47017 ohms

**Calibration Voltage:** 0.9097 volts

**Calibration Flow Rate:** 76.3

**Calibration Current:** 19.26 mA

**Zero Switch Setting:** 0

**TB3 Jumper Position:** + (POSITIVE)

**Span Switch Setting - Undamped:** 0

**Span Switch Setting - Damped:** 8

**Additional Information:**